

# BIOMARK LABORATORIES, INDIA

## CERTIFICATE OF ANALYSIS

**PRODUCT :** Oxacillin resistance screening agar base

**PRODUCT CODE NO.:** B1240

### QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Straw grey
APPEARANCE	Homogenous free flowing powder

### QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	7.2+/-0.2
COLOR (post autoclaving/heating)	Straw grey
CLARITY (post autoclaving/heating)	Slightly opalescent with slight precipitate

### QUALITY CONTROL TESTS : MICROBIOLOGICAL

**CULTURAL RESPONSE:** This is to certify that this lot passes and it confirms to the above mentioned tests and specifications. The information given Cultural characteristics observed with added Oxacillin Resistance Selective Supplement , after an incubation at 35-37°C for 18-24 hours.

here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of	
Organism Description	Growth
Staphylococcus aureus ATCC 29213	[Inhibited (susceptible to oxacillin)]
Staphylococcus aureus ATCC 38591	[Fair to good (presence of colony or haze of growth should be read as resistance).]

release.

This document has been produced electronically and is valid