

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT : Universal pre-enrichment broth

PRODUCT CODE NO.: B1275

QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Light yellow to beige
APPEARANCE	Homogenous free flowing powder

QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	6.3+/-0.2
COLOR (post autoclaving/heating)	Light amber
CLARITY (post autoclaving/heating)	Clear to slightly opalescent & may have slight precipitate

QUALITY CONTROL TESTS : MICROBIOLOGICAL

CULTURAL RESPONSE:

Cultural characteristics observed after an incubation at 35-37°C for 18-24 hours.

Organism Description	Inoculum (CFU)	Growth	
Listeria monocytogenes ATCC 19118	[50-100]	[Good]	
Salmonella Enteritidis ATCC 13076	[50-100]	[Good]	
Salmonella Typhimurium ATCC 14028	[50-100]	[Good]	

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.
Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** www.biomarklabs.com.

Rev: December 2023