

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT : Peptone tryptone water

PRODUCT CODE NO.: B1431

QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Cream to light yellow
APPEARANCE	Homogenous free flowing powder

QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	7.3+/-0.2
COLOR (post autoclaving/heating)	Light yellow
CLARITY (post autoclaving/heating)	Clear

QUALITY CONTROL TESTS : MICROBIOLOGICAL

CULTURAL RESPONSE:

Cultural characteristics observed after 24 hrs. at 35-37°C.

Organism Description	Inoculum	Growth	Indole test
Escherichia coli (25922)	[50-100]	[Luxuriant]	[[positive reaction, red ring at the interface of the medium on addition of Kovacs reagent]]
Klebsiella pneumoniae (23357)	[50-100]	[Luxuriant]	[[negative reaction, no red ring at the interface of the medium on addition of Kovacs reagent]]

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.
Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **WebSite:-** www.biomarklabs.co

Rev: December 2023