

# BIOMARK LABORATORIES, INDIA

## CERTIFICATE OF ANALYSIS

**PRODUCT :** Campylobacter Blood Free Medium Base Bolton - (mCCDA)

**PRODUCT CODE NO.:** B1457

### QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Grey to black
APPEARANCE	Homogenous free flowing powder

### QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	7.4+/-0.1
COLOR (post autoclaving/heating)	Black
CLARITY (post autoclaving/heating)	Opalescent gel

### QUALITY CONTROL TESTS : MICROBIOLOGICAL

#### CULTURAL RESPONSE:

Cultural characteristics observed with added Campylobacter Supplement V(FD067),after an incubation at 42°C for 24-48 hours.

Organism Description	Growth	Inoculum (CFU)	Recovery	Colour of
Campylobacter coli ATCC 33559	[good-luxuriant]	[50-100]	[>=50%]	[creamy-grey]
Campylobacter jejuni ATCC 29428 (00156*)	[good-luxurian]	[50-100]	[>=50%]	[Grey]
Escherichia coli ATCC 25922	[inhibited]	[>=104]	[0%]	[-]
Campylobacter laridis ATCC 35222	[good-luxuriant]	[50-100]	[>=50%]	[varying type]

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.