

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT : Acetamide broth (twin pack)

PRODUCT CODE NO.: B1500

QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Part A) Colourless ,Part B) Off white to white
APPEARANCE	Homogenous free flowing powder

QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	7.0+/-0.5
COLOR (post autoclaving/heating)	Part A) Colourless ,Part B) Off white to white
CLARITY (post autoclaving/heating)	clear solution

QUALITY CONTROL TESTS : MICROBIOLOGICAL

CULTURAL RESPONSE:

Cultural characteristics observed after an incubation at 35-37°C for 18-24 hours.

Organism Description	Inoculum (CFU)	Growth	Deamination
Pseudomonas aeruginosa ATCC27853	[50-100]	[good-luxuriant]	[Positive,yellow to brick red colour formation on addition of Nessler's reagent]
Stenotrophomonas maltophilia ATCC 13637	[50-100]	[good-luxuriant]	[negative,no colour formation on addition of Nessler's reagent]

This is to certify that this lot passes and it conforms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.