

# BIOMARK LABORATORIES, INDIA

## CERTIFICATE OF ANALYSIS

**PRODUCT :** Anaerobic broth

**PRODUCT CODE NO.:** B1805

### QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	cream to yellow
APPEARANCE	Homogenous free flowing powder

### QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	7.2+/-0.2
COLOR (post autoclaving/heating)	cream to yellow
CLARITY (post autoclaving/heating)	Clear solution

### QUALITY CONTROL TESTS : MICROBIOLOGICAL

#### CULTURAL RESPONSE:

Cultural characteristics observed under anaerobic conditions after an incubation at 35-37 °C for 48-72 hours.

Organism Description	Inoculum	Growth	Recovery
Clostridium butyricum ATCC 13732	[50-100]	[good-luxuriant]	[>=50%]
Clostridium perfringens ATCC 12924	[50-100]	[good-luxuriant]	[>=50%]
Clostridium sporogenes ATCC 11437	[50-100]	[good-luxuriant]	[>=50%]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.