

# BIOMARK LABORATORIES, INDIA

## CERTIFICATE OF ANALYSIS

**PRODUCT :** Phenol red maltose agar

**PRODUCT CODE NO.:** B290

### QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Light yellow to pink
APPEARANCE	Homogenous free flowing powder

### QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	7.4+/-0.2
COLOR (post autoclaving/heating)	Red
CLARITY (post autoclaving/heating)	Slightly opalescent

### QUALITY CONTROL TESTS : MICROBIOLOGICAL

#### CULTURAL RESPONSE:

Cultural characteristics observed after an incubation at 35 - 37°C for 18 - 24 hours

Organism Description	Inoculum(CFU)	Growth	Acid	Gas
Alcaligenes faecalis ATCC 8750	[50-100]	[luxuriant]	[Negative reaction ,no colour change]	[Negative reaction]
Escherichia coli ATCC 25922	[50-100]	[luxuriant]	[Positive reaction yellow colour]	[Positive reaction]
Klebsiella pneumoniae ATCC 13883	[50-100]	[luxuriant]	[Positive reaction yellow colour]	[Positive reaction]
Proteus vulgaris ATCC 13315	[50-100]	[luxuriant]	[Positive reaction yellow colour]	[Positive reaction]
Salmonella Typhimurium ATCC 14028	[50-100]	[luxuriant]	[Positive reaction yellow colour]	[Positive reaction]
Shigella flexneri ATCC 12022	[50-100]	[luxuriant]	[Positive reaction yellow colour]	[Negative reaction ]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.