

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT : Wilkins chalgren anaerobic broth base

PRODUCT CODE NO.: B358

QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Cream to yellow
APPEARANCE	Homogenous free flowing powder

QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	7.1+/-0.2
COLOR (post autoclaving/heating)	Medium amber
CLARITY (post autoclaving/heating)	Clear

QUALITY CONTROL TESTS : MICROBIOLOGICAL

CULTURAL RESPONSE:

Cultural characteristics observed with added Non-Spore Anaerobic Supplement (BF100) or G.N.Spore Anaerobic Supplement (BF101) Under anaerobic conditions, after an incubation at 35-37°C of 48 hours.

Organism Description	Inoculum (CFU)	Growth
Bacteroides fragilis ATCC 25285	[50-100]	[Luxuriant]
Clostridium perfringens ATCC 12924	[50-100]	[Luxuriant]
Prevotella melaninogenicus ATCC 15930	[50-100]	[Luxuriant]
Escherichia coli ATCC 25922	[>=10 ³]	[Inhibited]

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.