

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT: SABHI AGAR BASE

PRODUCT CODE NO: B376

QUALITY CONTROL TESTS: DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOUR	Yellow
APPEARANCE	Homogeneous Free Flowing powder

QUALITY CONTROL TESTS: REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH (post autoclaving/heating)	7.0 ± 0.2
COLOR (post autoclaving/heating)	a) Basal medium : Yellow b) With the addition of 10% v/v blood : Cherry red
CLARITY (post autoclaving/heating)	a) Clear b) Opaque

QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE:

Cultural characteristics observed after 40 - 48 hrs at 25 - 30°C.

Organism Description	GROWTH W/O BLOOD*	GROWTH W/ BLOOD*
Candida albicans (ATCC10231)	Good - luxuriant	Luxuriant
Saccharomyces cerevisiae (ATCC9763)	Good - luxuriant	Luxuriant
Saccharomyces uvarum (ATCC9080)	Good - luxuriant	Luxuriant
Aspergillus niger (ATCC16404)	Good	Good
Escherichia coli (ATCC25922)	Inhibited	Inhibited
Staphylococcus aureus (ATCC25923)	Inhibited	Inhibited

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.
Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** www.biomarklabs.com

Rev: January 2026