

CERTIFICATE OF ANALYSIS

PRODUCT : Alkaline peptone water

PRODUCT CODE NO.: B398

QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Cream to Light yellow
APPEARANCE	Homogenous free flowing powder

QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	8.4+/-0.2
COLOR (post autoclaving/heating)	Pale yellow to light yellow
CLARITY (post autoclaving/heating)	clear solution

QUALITY CONTROL TESTS : MICROBIOLOGICAL

CULTURAL RESPONSE:

Cultural characteristics observed after an incubation at 35-37°C for 18-24 hours.

Organism Description	Inoculum	Growth
Vibrio parahaemolyticus (17802)	50-100	luxuriant
Vibrio cholerae (15748)	50-100	luxuriant

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.