

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT : Pseudomonas asparagine broth

PRODUCT CODE NO.: B414

QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

| PARAMETERS | SPECIFICATIONS/STANDARD VALUES |
|------------|--------------------------------|
| COLOR | Offwhite to white |
| APPEARANCE | Homogenous free flowing powder |

QUALITY CONTROL TESTS : REHYDRATED MEDIUM

| PARAMETERS | SPECIFICATIONS/STANDARD VALUES |
|------------------------------------|--|
| pH(post autoclaving/heating) | 7.0+/-0.2 |
| COLOR (post autoclaving/heating) | Colourless to very light yellow |
| CLARITY (post autoclaving/heating) | clear solution with slight precipitate |

QUALITY CONTROL TESTS : MICROBIOLOGICAL

CULTURAL RESPONSE:

Cultural characteristics observed after 20 -24 hrs.at 35-37°C.

| Organism Description | Microorganisms (ATCC) | Growth | Inoculum (CFU) |
|--------------------------------|----------------------------------|--------------|----------------|
| Pseudomonas aeruginosa (27853) | [Pseudomonas aeruginosa (27853)] | [luxuriant] | [50-100] |

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.