

# BIOMARK LABORATORIES, INDIA

## CERTIFICATE OF ANALYSIS

**PRODUCT :** Dextrose proteose peptone agar

**PRODUCT CODE NO.:** B492

### QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Cream to yellow
APPEARANCE	Homogenous free flowing powder

### QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	7.4+/-0.2
COLOR (post autoclaving/heating)	Reddish brown after addition of blood & tellurite solution
CLARITY (post autoclaving/heating)	Opaque

### QUALITY CONTROL TESTS : MICROBIOLOGICAL

#### CULTURAL RESPONSE:

Cultural characteristics observed after 24 – 48 hrs.at 35-37°C.

Organism Description	Organism	Inoculum (CFU)	Growth	Recovery
corynebacterium diphtheriae (ATCC 11913)	[Complies]	[50-100]	[Good -Luxuriant]	[>=50%]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.