

CERTIFICATE OF ANALYSIS

QA/F/01 Rev no. 03

PRODUCT :	Tryptone bile agar
PRODUCT CODE NO.:	B788

QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Cream to yellow
APPEARANCE	Homogenous free flowing powder

QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	7.2+/-0.2
COLOR (post autoclaving/heating)	Cream to yellow
CLARITY (post autoclaving/heating)	Clear to slightly opalescent

QUALITY CONTROL TESTS : MICROBIOLOGICAL

CULTURAL RESPONSE:

Cultural characteristics observed after an incubation at 35-37°C for 24-48 hours.

Organism Description	Inoculum(CFU)	Growth	Recovery
Enterobacter aerogenes ATCC 13048	[>=10 ³]	[inhibited]	[0%]
Escherichia coli ATCC 25922	[50-100]	[good-luxuriant]	[>=50%]