

## CERTIFICATE OF ANALYSIS

**PRODUCT :** Trytpose agar

**PRODUCT CODE NO.:** B804

### QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Cream to yellow
APPEARANCE	Homogenous free flowing powder

### QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	7.2+/-0.2
COLOR (post autoclaving/heating)	a)Basal medium-Yellow.b)With addition of 5% V/v sterile defibrinated blood-Cherry red
CLARITY (post autoclaving/heating)	a)Clear to slightly opalescent.b)Opaque

### QUALITY CONTROL TESTS : MICROBIOLOGICAL

#### CULTURAL RESPONSE:

Cultural characteristics observeds after an incubation at 35-37°C for 48-72 hours with added 5% v/v sterile defifibrinated blood in presence of 10% Carbon dioxide.

Organism Description	Growth
Brucella melitensis ATCC4309	[good-luxuriant]
Brucella suis ATCC 4314	[good-luxuriant]
Streptococcus pneumoniae ATCC 6303	[good-luxuriant]
Streptococcus pyogenes ATCC19615	[good-luxuriant]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.