

## CERTIFICATE OF ANALYSIS

**PRODUCT :** Giemsa stain

**PRODUCT CODE NO.:** BA011

**QUALITY CONTROL:** Physical

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Colour	Dark blue
Appearance	Liquid
Clarity	Clear

**QUALITY CONTROL: MICROSCOPY (STAINING)**

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Cytoplasm	clear blue
Nuclear and basophilic cytoplasmic components	Blue
Neutrophilic granules	Lilac
Eosinophilic granules	Orange
Mast cell granules	Deep blue violet
Nucleoli	blue violet
Red blood cells	pink
cytoplasm of mature monocytes	gray blue

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.