

CERTIFICATE OF ANALYSIS

PRODUCT: THYMOL BLUE INDICATOR PRODUCT CODE NO: BA056
--

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOUR	Blue
APPEARANCE	Liquid
CLARITY	Clear

QUALITY CONTROL: MICROSCOPY (STAINING) NA

PARAMETERS	SPECIFICATION
Reaction	pH Range : 1.2 – 2.8 (Acidic) 8.0 – 9.6 (Alkaline) Colour change: Red to yellow(Acidic) Yellow to blue(Alkaline)

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.