

CERTIFICATE OF ANALYSIS

PRODUCT : Barrit reagent b**PRODUCT CODE NO.:** BA085**QUALITY CONTROL:** Physical

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Colour	Colourless
Appearance	Liquid
Clarity	Clear

QUALITY CONTROL: MICROSCOPY (STAINING)

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Cultural response	Biochemical identification was carried out by adding 0.2 ml of Barritt Reagent (Part A) and 0.2 ml of Barritt Reagent (Part B)
Organism :	Growth & VP Test
Enterobacter aerogenes ATCC 13048	Luxuriant Positive (Cherry red colour formation)
Escherichia coli ATCC 25922	Luxuriant Negative (No red colour formation)
Klebsiella pneumoniae ATCC 13883	Luxuriant Positive (Cherry red colour formation)

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.