

## CERTIFICATE OF ANALYSIS

**PRODUCT :** Gordon - mclead reagent (oxidase reagent)**PRODUCT CODE NO.:** BA101**QUALITY CONTROL:** Physical

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Colour	Light yellow colour
Appearance	Liquid
Clarity	Clear

**QUALITY CONTROL: MICROSCOPY (STAINING)**

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Organism	Oxidase Reaction
Escherichia coli (ATCC 25922)	Negative(No change in colour)
Pseudomonas aeruginosa (ATCC 27853)	Positive(development of purple-blue colour)

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.