

## **CERTIFICATE OF ANALYSIS**

**PRODUCT: Cefepime 30 mcg**

**PRODUCT CODE NO.: BD014**

### **QUALITY CONTROL: PHYSICAL**

<b>PARAMETERS</b>	<b>SPECIFICATIONS/STANDARD VALUES</b>
<b>APPEARANCE</b>	Filter paper discs of 6mm diameter With <b>CPM</b> printed on discs.

### **QUALITY CONTROL: MICROBIOLOGICAL**

Average diameter of zone of inhibition in mm observed on Mueller Hinton Agar Plates (B263) at 35-37°C for 18-24 hours.

<b>Micro-organism with ATCC number</b>	<b>Standard diameter of zone of inhibition in mm</b>
	Quality control limit in mm
E.coli ATCC 25922	31-37
S.aureus ATCC 25923	23-29
Ps.aeruginosa ATCC 27853	24-30

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.