BIOMARK Laboratories – INDIA

www.biomarklabs.com

CERTIFICATE OF ANALYSIS

PRODUCT: Moxifloxacin 5mcg

PRODUCT CODE NO.: BD048

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES	
APPEARANCE	Filter paper discs of 6mm diameter with MX printed on discs	

QUALITY CONTROL: MICROBIOLOGICAL

Average diameter of zone of inhibition in mm observed on Mueller Hinton Agar Plates (B263) at 35-37°C for 18-24 hours.

Antimicrobial Disc with	Test Organisms (ATCC)	Standard diameter of zone of inhibition in mm		
concentration		Resistant	Intermediate	Sensitive
Moxifloxacin (5 mcg)	E.coli (25922)	≤ 20	21-23	≥24
	S.aureus (25923)	≤ 20	21-23	≥24
	P.aeruginosa (27853)	≤ 20	21-23	≥24