

CERTIFICATE OF ANALYSIS

PRODUCT: PIPERACILLIN/TAZOBACTAM (PT 100/10 mcg)

PRODUCT CODE NO.: BD061

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
APPEARANCE	Filter paper discs of 6mm diameter with PT printed on discs

QUALITY CONTROL: MICROBIOLOGICAL

Average diameter of zone of inhibition in mm observed on Mueller Hinton Agar Plates (B263) at 35-37°C for 18-24 hours as per CLSI / NON CLSI standards.

, micro-organism with ATCC number	Diameter of zone of inhibition
	Quality control limit in mm
(PT 100/10 mcg)	
EC (ATCC 25922)	24-30
SA (ATCC 25923)	27-36
PS (ATCC 27853)	25-33

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.