BIOMARK Laboratories – INDIA

www.biomarklabs.com

CERTIFICATE OF ANALYSIS

PRODUCT: Furazolidone 100 mcg

PRODUCT CODE NO.: BD077

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES	
APPEARANCE	Filter paper discs of 6mm diameter with FX printed on discs	

QUALITY CONTROL: MICROBIOLOGICAL

Average diameter of zone of inhibition in mm observed on Mueller Hinton Agar Plates (B263) at 35-37°C for 18-24 hours.

Antimicrobial Disc with	Test Organism (ATCC)	Standard diameter of zone of inhibition in mm		
concentration		Resistant	Intermediate	Sensitive
Furazolidone 100 mcg	E.coli (25922)	-	-	≥ 20