

CERTIFICATE OF ANALYSIS

PRODUCT : CEFTAZIDIME-CLAVULARIC ACID (CAC 30/10 mcg)**PRODUCT CODE NO.: BD081****QUALITY CONTROL: PHYSICAL**

| PARAMETERS | SPECIFICATIONS/STANDARD VALUES |
|------------|--|
| APPEARANCE | White coloured discs of 6mm diameter with CAC printed |

QUALITY CONTROL: MICROBIOLOGICAL

Average diameter of zone of inhibition in mm observed on Mueller Hinton Agar Plates (B263) at 35-37°C for 18-24 hours as per CLSI / NON CLSI standards.

| micro-organism with ATCC number | Diameter of zone of inhibition |
|---------------------------------|--------------------------------|
| | Quality control limit in mm |
| EC (ATCC 25922) | 27-34 |
| SA (ATCC 25923) | 18-22 |
| PS (ATCC 27853) | -- |

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.