## BIOMARK Laboratories-INDIA www.biomarklabs.com

## **CERTIFICATE OF ANALYSIS**

PRODUCT: KOVAC'S REAGENT STRIPS

PRODUCT CODE NO.: BDD024

## **QUALITY CONTROL: PHYSICAL**

PARAMETERS	SPECIFICATIONS/ STANDARD VALUES	
APPEARANCE	Filter paper strips of 60 cm x 5 mm	

## **QUALITY CONTROL: MICROBIOLOGICAL**

Cultural response observed in Peptone Water (B035) at 35-37°C for 18-24 hours by inserting the Kovac's reagent strip between the plug and inner wall of the tube.

Organism(ATCC)	Indole Production
Escherichia coli (25992)	Positive (Pink coloration of lower part of the strip)
Enterobacter aerogenes (13048)	Negative (No color change).