

BIOMARK LABORATORIES

CERTIFICATE OF ANALYSIS

PRODUCT: KOVAC'S REAGENT STRIPS

PRODUCT CODE NO.: BDD024

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
APPEARANCE	Paper strips of size 60mmx5 mm

QUALITY CONTROL: MICROBIOLOGICAL

Cultural response observed in Peptone Water (B035) at 35-37°C for 18-24 hours by inserting the Kovac's reagent strip between the plug and inner wall of the tube.

ORGANISM DESCRIPTION (ATCC)	Indole Production
Escherichia coli (25992)	Positive (Pink coloration of lower part of the strip)
Enterobacter aerogenes (13048)	Negative (No color change).

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.
Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** www.biomarklabs.com.

Rev: January 2025