

BIOMARK LABORATORIES

CERTIFICATE OF ANALYSIS

PRODUCT: LEAD ACETATE PAPER STRIPS

PRODUCT CODE NO.: BDD025

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
APPEARANCE	Paper strips of size 60mmx5 mm.

QUALITY CONTROL: MICROBIOLOGICAL

Cultural response observed in Peptone Water (B035) at 35-37°C for 18-24 hours by inserting the lead acetate strip between the plug and inner wall of the tube.

ORGANISM DESCRIPTION (ATCC)	H2S production
Escherichia coli (25992)	Negative (No blackening of the lower part of the strip)
Salmonella typhimurium (14028)	Positive (Blackening of the lower part of the strip)

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.
Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** www.biomarklabs.com.

Rev: January 2025