

BIOMARK LABORATORIES

CERTIFICATE OF ANALYSIS

PRODUCT: OXIDASE DISCS

PRODUCT CODE NO.: BDD027

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
APPEARANCE	Paper discs of 6 mm diameter

QUALITY CONTROL: MICROBIOLOGICAL

The reaction is observed within 5-10 seconds at 25-30°C

ORGANISM DESCRIPTION (ATCC)	Reaction	Colour
Pseudomonas aeruginosa (27853)	Positive	Deep purple blue
Neisseria gonorrhoeae (19424)	Positive	Deep purple blue
Staphylococcus aureus(25923)	Negative	No colour change
Escherichia coli(25992)	Negative	No colour change

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.

Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** www.biomarklabs.com.

Rev: January 2025