BIOMARK LABORATORIES

CERTIFICATE OF ANALYSIS

PRODUCT CODE NO.: BDD027

QUALITY CONTROL: PHYSICAL

| PARAMETERS | SPECIFICATIONS/STANDARD VALUES |
|------------|-----------------------------------|
| APPEARANCE | Filter paper discs of 6mmdiameter |

QUALITY CONTROL: MICROBIOLOGICAL

Typical oxidase reaction given by 18-48-hourculture within 5-10 seconds at 25-30°C.

| Organism(ATCC) | Reaction |
|--------------------------------|--|
| Pseudomonas aeruginosa (27853) | Positive: Deep purplish blue colouration of disc |
| Staphylococcus aureus(25923) | Negative: No colour change |
| Escherichia coli (259222) | Negative: No colour change |

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.