

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT : Ttc solution 1% (10 ml per vial)

PRODUCT CODE NO: BF044

QUALITY CONTROL:

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Colour	Colourless
Appearance	Liquid
Clarity	Clear

QUALITY CONTROL : MICROBIOLOGICAL

CULTURAL RESPONSE:

Cultural characteristics observed with added BF067 and BF044, after an incubation at 35-37°C for 48-72 hours.

Orgm Desc	INOCULUM(CFU)	GROWTH	RECOVERY	COLOUR OF COLONEY
Enterococcus aerogenes (13048)	$\geq 10^3$	inhibited	0%	-
Enterococcus faecalis (29212)	50-100	good-luxuriant	$\geq 50\%$	red-maroon
Escherichia coli (25922)	$\geq 10^3$	inhibited	0%	-

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.