

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT : Urea-40% (5 ml per vial)

PRODUCT CODE NO: BF048

QUALITY CONTROL:

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Colour	Colourless
Appearance	Liquid
Clarity	Clear

QUALITY CONTROL : MICROBIOLOGICAL

CULTURAL RESPONSE:

Cultural characteristics observed after (24) hours at (37)°C when added to Urea agar base(B347)

Orgm Desc	Inoculum	Growth	Urease
Salmonella typhimurium ATCC 14028	50-100	luxuriant	negative reaction,no change
Enterobacter aerogenes	50-100	luxuriant	negative reaction,no change
Escherichia coli ATCC 25922	50-100	luxuriant	negative reaction,no change
Proteus vulgaris ATCC13315	50-100	luxuriant	positive reaction,cerise colour

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.

Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** www.biomarklabs.com.

Rev: December 2023