

CERTIFICATE OF ANALYSIS

PRODUCT :	V.c.n.supplement
PRODUCT CODE NO:	BF059

QUALITY CONTROL:

PARAMETERS	SPECIFICATIONS/STANDARD VALUES	
Colour	White	
Appearance	Powder	
Clarity	NA	

QUALITY CONTROL: MICROBIOLOGICAL CULTURAL RESPONSE:

Cultural characteristics observed after 24 hours at 35-37°C when added to GC agar base (B1049) alongwith 2% Haemoglobin solution and GC supplement (BF058).

Orgm Desc	Inoculum(CFU)	Growth	Recovery	
Haemophilus influenzae ATCC19418	50-100	Good-luxuriant	>=50%	
Neisseria gonorrhoeae ATCC19424	50-100	Good-luxuriant (with added Antibiotic supplements)	>=50%	
Neisseria meningitidis ATCC13090	50-100	Good-luxuriant (with added Antibiotic supplements)	>=50%	
Streptococcus pyogenes ATCC19615	50-100	Good-luxuriant	>=50%	
Streptococcus pneumoniae ATCC6303	50-100	Good-luxuriant	>=50%	