BIOMARK LABORATORIES

CERTIFICATE OF ANALYSIS

PRODUCT: Potato dextrose agar

PRODUCT CODE NO.: PP020-50

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/ STANDARD VALUES		
Colour	light yellow		
Appearance	Sterile PDA in 90X15 mm dis. plates.		
Clarity	Slightly opalascent		
Sterility after 48 hrs	Sterile		

QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE:

Cultural characteristics observed after (48) hours at (25-30)oC

Orgm Desc	Inoculum(CFU)	Growth	Ascospore formation	Recovery	
Aspergillus brasiliensis ATCC 16404	50-100	luxuriant	Negative	-	
Candida albicans ATCC10231	50-100	luxuriant	negative	>=70%	
Saccharomyces cerevisiae ATCC 9763	50-100	luxuriant	Positive	>=70%	

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.