

CERTIFICATE OF ANALYSIS

PRODUCT : R- 2A agar

PRODUCT CODE NO.: PP023-50

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/ STANDARD VALUES
Colour	Light yellow coloured medium
Appearance	Sterile R- 2A in 90X15 mm disposable plates
Clarity	Slightly opalescent
Sterility after 48hrs	Sterile

QUALITY CONTROL TESTS: MICROBIOLOGICAL

CULTURAL RESPONSE:

Cultural characteristics observed *by using standard ATCC cultures after an incubation at 35-37°C for 24-72 hours. (*-In case of water samples from fields it is suggested to incubate further for up to 7 days to ascertain the absence of organisms)

Orgm Desc	Inoculum	Growth
Candida albicans ATCC 10231	50-100	good-luxuriant
Enterococcus faecalis ATCC 29212	50-100	good-luxuriant
Escherichia coli ATCC 25922	50-100	good-luxuriant
Salmonella Enteritidis ATCC 13076	50-100	good-luxuriant
Salmonella typhi ATCC 6539	50-100	good-luxuriant

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.