CERTIFICATE OF ANALYSIS

PRODUCT : Sheep blood agar

PRODUCT CODE NO.: PP028-20

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/ STANDARD VALUES		
Colour	Bright Red colour		
Appearance	Sterile Sheep Blood Agar in 85mm disposable plates.		
Clarity	Opaque		
Sterility after 24hrs	Sterility		

QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE:

Cultural characteristics observed after (18-48) hours at (35-37)oC.

Orgm Desc	Inoculum (CFU)	Growth	Recovery	Haemolysis
Streptococcus pneumoniae ATCC 6303	50-100	luxuriant	>=70%	alpha
Streptococcus pyogenes ATCC 19615	50-100	luxuriant	>=70%	beta

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA. Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, Website:- www.biomarklabs.com.