

CERTIFICATE OF ANALYSIS

PRODUCT :	Wilson and Blair's BBS Agar Medium
PRODUCT CODE NO.:	PP036-20

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/ STANDARD VALUES
Colour	Greenish yellow
Appearance	Sterile Wilson and Blair's BBS Agar in 85mm disposable plates.
Clarity	Opaque
Sterility after 24hrs	Sterile

QUALITY CONTROL TESTS: MICROBIOLOGICAL

CULTURAL RESPONSE:

Cultural characteristics observed after (24-48) hours at (35-37)oC.

Orgm Desc	Inoculum (CFU)	Growth	Recovery
Escherichia coli ATCC 25922	$\geq 10^3$	inhibited	0%
Proteus mirabilis ATCC 25933	50-100	Luxuriant	$\geq 50\%$
Salmonella Typhi ATCC 6539	50-100	Luxuriant	$\geq 50\%$
Salmonella Typhimurium ATCC 14028	50-100	Luxuriant	$\geq 50\%$

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.