

CERTIFICATE OF ANALYSIS

PRODUCT : Non-nutrient Agar 1.5%**PRODUCT CODE NO.:** PP039-10**QUALITY CONTROL: PHYSICAL**

PARAMETERS	SPECIFICATIONS/ STANDARD VALUES
Colour	Straw
Appearance	Sterile Non Nutrient Agar1.5% in 90X15 mm disposable plates
Clarity	Clear
Sterility after 48 hrs	Sterile

QUALITY CONTROL TESTS: MICROBIOLOGICAL**CULTURAL RESPONSE:**

Cultural characteristics observed after (3-24) hours up to 10days at (30-37)°C for Naegleria fowleri & Acanthamoeba spp.

Orgm Desc	Growth
Acanthamoeba castellanii ATCC30010	Luxuriant -double wall cyst
Naegleria fowleri ATCC30215	Luxuriant
Escherichia coli ATCC 10536	Inhibition
Escherichia coli ATCC 25922	Inhibition

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

