#### **BIOMARK Laboratories-INDIA**

## www.biomarklabs.com

## **TECHNICAL SHEET**

B263 M	MUELLER HINTON AGAR				
Formula					
Ingredients :		gms/lit.			
Meat, infusion from		300.00			
Casein acid hydrolysate		17.50			
Starch		1.50			
Agar		17.00			
# Equivalent to beef, in	nfusion form				
Final pH (at 25°C):	7.3+ 0.1				
Directions :	<del></del>				

Suspend 38.0 grams in 1000 ml purified/ distilled water. Heat to boiling to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool to 45-50°C. Mix well and pour into sterile Petri plates

Note: The performance of this batch has been tested and standardized as per the current CLSI (formerly, NCCLS) document M6-protocols for Evaluating Dehydrated Mueller Hinton Agar.

## Principle:

Meat, infusion from and casein acid hydrolysate provide nitrogenous compounds, carbon, sulphur and other essential nutrients. Starch is added to absorb any toxic substances present in the medium. Different factors influence the disc diffusion susceptibility tests as, inoculum concentration, agar depth, disc potency, medium pH and beta – lactamase production by test organisms. Agar is the solidifying agent.

Torganisms. Agains the soliditying agent.						
QC Tests - (I)Dehydrated Medium						
Colour:	Cream to yellow					
Appearance :	Homogeneous Free Flowing powder					
(II)Rehydrated medium						
pH (post autoclaving/heating) :	$7.3 \pm 0.1$					
Colour (post autoclaving/heating):	Light amber					
Clarity (post autoclaving/heating):	Sligthly opalescent					
(III) Q.C. Test Microbiological						
Cultural characteristics observed after 18- 24	Cultural characteristics observed after 18- 24 hours at 36-37°C.					
MICROORGANISM (ATCC )	GROWTH					
Escherichia coli (25922)	Luxuriant					
Pseudomonas aeruginosa (27853)	Luxuriant					
Staphylococcus aureus (25923)	Luxuriant					
Enterococcus faecalis(29212)	Luxuriant					
Escherichia coli (35218)	Luxuriant					
Staphylococcus aureus subsp. aureus (43300)	Luxuriant					
Precautions: 1. For Laboratory Use.						
2. Follow proper, established laborated infectious materials.	pratory procedures in handling and disposing of					

Refer disclaimer Overleaf

Page 01 of 02

## **BIOMARK Laboratories-INDIA**

# www.biomarklabs.com

## **TECHNICAL SHEET**

Limitations :	1. Since the nutritional requirements of organisms vary, some strains may be									
	encountered that fail to grow or grow poorly on this medium.									
	2. Numerous factors can affect results ; inoculum size, rate of growth, medium									
	formulation of pH, length of incubation and incubation environment, disk content and									
	drug diffusion rate, and measurement of endpoints. Therefore, strict adherence to									
		red to ensure reliable								
		susceptibility testing								
		result from the prolo								
		ing excessive amoun								
	inhibitory effects of sulfonamides and trimethoprim, causing zones of growth inhibition to									
	be smaller or less distinct.									
	5. Variation in the concentration of divalent cations, primarily calcium and magnesium,									
	affects results of aminoglycoside, tetracycline, and colistin tests with P. aeruginosa									
	isolates. A cation content that is too high reduces zones sizes, whereas a cation content									
	that is too low has the opposite effect.									
	6. When Mueller Hinton Medium is supplemented with blood, the Zone of inhibition for									
	oxacillin and methicillin may be 2 to 3 mm smaller than those obtained with									
		unsupplemented agar. Conversely, sheep blood may markedly increase the zone diameters of some cephalosporins when they are tested against enterococci. Sheep blood								
		may cause indistinct zones or a film of growth within the zones of inhibition around sulfonamide and trimethoprim disks.								
	7. Mueller Hinton Medium deeper than 4 mm may cause false – resistant results, and									
	agar less than 4 mm deep may be associated with a false –suceptibility report.									
		8. A pH outside the range of $7.3 \pm 0.1$ may adversely affect susceptibility test results. If								
	the pH is too low, aminoglycosides and macrolides will appear to lose potency; others may appear to have excessive activity. The opposite effects are possible if the pH is too									
	high.		•	• •	•	'				
	9. When Mueller Hinton Medium is inoculated, no droplets of moisture should be visible									
	on the surface or on the petri dish cover.									
	10. Mueller Hinton Medium should be inoculated within 15 minutes after the inoculum									
	suspension has been adjusted.									
	11. The zone of inhibition diameters of some drugs, such as the aminoglycosides, macrolides, and tetracyclines, are significantly altered by $CO_2$ Plates should not be incubated in increased $CO_2$ .									
		is recommended for								
Use:		or determination of s		ibility of	f microorganisms	to antimicrobial				
		from clinical samples.								
Storage :	Dehydrated medium -below 30°C Prepared medium - Between 2 to 8°C.									
Packing:	500 gm. bottle									
Product profile:	Reconstitution	Quantity on	pH (	25°C)	Supplement	Sterilization				
		Preparation (500g)								
B263	38g/l	13.157L	7.3	± 0.1	NIL	121°C / 15 minutes				

#### Disclaimer:

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications.

The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.