BIOMARK Laboratories-INDIA www.biomarklabs.com

TECHNICAL SHEET

B638	MOTILITY NITRATE MEDIUM, BUFFERED						
Formula							
Ingredients:	gms/lit.						
Peptic digest of animal ti	ssue 5.00						
Meat Extract B#	3.00						
Galactose	5.00						
Potassium nitrate	1.00						
Disodium phosphate	2.50						
Agar	3.00						
#- Equivalent to Beef ext	tract						
Final pH (at 25°C):	7.3 <u>+</u> 0.2						
Directions :		·					
C	000 and distilled containing Fool absence Heat	. And the officers have although the con-					

Suspend 19.5 gms. in 1000 ml distilled water containing 5ml. glycerol. Heat to boiling to dissolve the medium completely. Dispense in test tubes to make them half full. Sterilize by autoclaving at 15 lbs pressure (121° C) for 15 minutes. Cool quickly in cool running water and allow the tubed medium to solidify in an upright position.

Principle:

Peptic digest of animal tissue and Meat Extract B galactose provide essential nutrients for growth. Potassium nitrate is the substrate for nitrate reduction which is detected with the help of two reagents, viz. sulfanilic acid (1 gm in 125 ml 5 N acetic acid) and N – (1-naphthyl) ethylene diamine dihydrochloride (0.25 gm in 200 ml 5N acetic acid). The presence of less quantity of agar in the medium makes it semisolid which allows detection of motility.

	moona winch a			•					
QC Tests - (I)Dehydrated Medium									
	Colour :			Light to medium yellow					
	Appearance :			Homogeneous Free Flowing powder					
(II	(II)Rehydrated medium								
	pH (post autoclaving/heating) :			7.3 ± 0.2					
	Colour (post autoclaving/heating):			Light amber					
	Clarity (post autoclaving/heating):			Clear to slightly opalascent gel forms in tubes as butt					
(III) Q.C. Test Microbiological									
	Cultural characteristics observed after 2			4- 48 hours at 35-37°C.					
	MICROORGANISM (ATCC)		MOTILIT	Υ	NITRATE REDUCTION				
	Clostridium perfringens (12924)			-			+		
	Clostridium absonum (27555)		W	W		±			
	Key: + = red-	absent							
	- = grow	ly motile							
Precautions: 1. For Laboratory Use.									
		2. Follow proper, established laboratory procedures in handling and disposing of							
	infectious materials.								
Limitations :		1. Since the nutritional requirements of organisms vary, some strains may be							
			grow or grow poorly on this medium.						
Use :		For isolation and detection of Clostridium perfringens on the basis of motility and nitrate							
test.									
				low 30°C Prepared medium- Between 2 to 8°C.					
Packing: 500 gm. bottle					1				
Product profile:		Reconstitution			рН ((25°C)	Supplement	Sterilization	
				ion (500g)					
B638		19.5 g/l	2!	5.64L	7.3	± 0.2	glycerol	121°C / 15 minutes	

Refer disclaimer Overleaf

Rev: December 2020

BIOMARK Laboratories-INDIA www.biomarklabs.com TECHNICAL SHEET

Disclaimer:

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications.

The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

Rev: December 2020