

B917	ADAMS AGAR					
Formula						
Ingredient:		gms/lit.				
Dextrose		0.40				
Sodium acetate		2.30				
Agar		20.00				
Final pH (at 25°C): Self						
Directions:						
Suspend 22.7 gms. in 1000 ml distilled water. Heat to boiling to dissolve the medium completely. Dispense in test tubes. Sterilize by autoclaving at 108-112°C for 15 minutes. Allow the tubes to solidify in slanted position.						
Principle:						
Dextrose in the medium stimulates sporulation. Acetate and dextrose are used as a carbon sources.						
QC Tests – (I) Dehydrated Medium						
Colour:		Off white-light yellow				
Appearance:		Homogeneous Free Flowing powder				
(II) Rehydrated medium						
pH (post autoclaving/heating) :		Self				
Colour (post autoclaving/heating):		Yellow				
Clarity (post autoclaving/heating):		Opalescent				
(III) Q.C. Test Microbiological						
Cultural characteristics observed after 18 –48 hrs. at 30°C.						
MICROORGANISM (ATCC)		GROWTH	SPORULATION			
Saccharomyces cerevisiae (9763)		luxuriant	positive			
Aspergillus brasiliensis (16404)		luxuriant	negative			
Candida albicans (10231)		luxuriant	negative			
Penicillium notatum (10108)		luxuriant	negative			
Precautions :		1. For Laboratory Use. 2. Follow proper, established laboratory procedures in handling and disposing of infectious materials.				
Limitations :		1. Since the nutritional requirements of organisms vary, some strains may be encountered that fail to grow or grow poorly on this medium.				
Use:		It is used for examining sporulation in yeasts.				
Storage:		Dehydrated medium- below 30°C Prepared medium– Between 2 to 8°C.				
Packing:		500 gm bottle				
Product profile:		Reconstitution	Quantity on Preparation (500g)	pH (25°C)	Supplement	Sterilization
B917	22.7 g/l	22.03 L	Self	Nil	108-112°C / 15 minutes	

Disclaimer:

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARK LABORATORIES publications.

The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.