BIOMARK Laboratories-INDIA www.biomarklabs.com

TECHNICAL SHEET

BA001

CARBOL FUCHSIN (ZN, DILUTE)

Formula

Reagents:

Carbol Fuchsin 1.00 ml Distilled water 19.00 ml

Directions:

- 1. Prepare a smear on clean, dry glass slide.
- 2. Air dry and fix it with gentle heat.
- 3. Flood slide with carbol fuchsin (Zn, dilute) for 2-3 minutes.
- 4. Wash with water.
- 5. Dry and observe under oil immersion lens.

Precautions:

- 1. For Laboratory Use.
- 2. Follow proper, established laboratory procedures in handling and disposing of infectious materials.

Use: Monochrome staining.

Quality Control:

Appearance: Magenta coloured clear solution

Microscopy: Bacteria are stained red against a colourless background.

Storage: Below 30°C.

Packing:100 ml

Disclaimer:

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications.

The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

Page 01 of 01