

BD010 CEFACLOR (CFC 30 mcg)

SPECIFICATIONS:

COLOUR: White discs of 6 mm diameter with abbreviated name "CFC".

DIRECTIONS:

1. Prepare sterile plates of Mueller Hinton Agar (B263).
2. Prepare inoculum as per Bauer-Kirby Method-make a pure inoculum of culture. Inoculate pure single colony of test organism – E. coli (ATCC 25922), S. aureus (ATCC 25923) and P. aeruginosa (ATCC 27853) in 5ml Pre tested sterile SCDM (B040) in a test tube & incubate at 36-37°C for 2 hours. Compare the inoculum turbidity with that of standard 0.5McFARLAND STANDARD (BA 132).
3. Dip a sterile swab into the standardized inoculum and press against the wall of the test tube to remove excess liquid.
4. Swab the plates by rotating in all directions till the plate becomes dry.
5. Place the antimicrobial ring on the swabbed plates aseptically with the help of alcohol dipped flame sterilized forceps. Press the ring properly on the plates.
6. Incubate the plates at 36-37°C for 18- 24 hours.
7. Measure the zones showing complete inhibition and record the diameters of the zones to the nearest millimetre using a calibrated instrument like zone scales.

RESULTS:

Measure the zone diameters in mm observed in the plates and compare them with the CLSI zone chart.

Antimicrobial agent, micro-organism with ATCC number	Diameter of zone of inhibition			
	Interpretative criteria			Quality control limit in mm
	Sensitive	Intermediate	Resistant	
	mm or more	mm	mm or less	
Cefaclor (30 mcg) CFC				
EC (ATCC 25922)	18	15-17	14	23-27
SA (ATCC 25923)	18	15-17	14	27-31
PS (ATCC 27853)	--	--	--	--

USE: It is used for antimicrobial sensitivity for relevant test organisms for laboratory diagnosis.

Precautions :For Laboratory Use.

Storage: -20 to +8°C

Packing:50 discs in vial

BIOMARK Laboratories-INDIA

www.biomarklabs.com

TECHNICAL SHEET

Disclaimer:

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications.

The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.