

BDD022

BACITRACIN DISCS

SPECIFICATIONS:

APPEARANCE: Paper disc of 6 mm diameter

DIRECTIONS:

Prepare Tryptose Blood agar plates (B342) and inoculate with suspect organism by surface spreading to obtain confluent growth. Place the Bacitracin disc aseptically onto the inoculated surface and press gently. Invert the plates and incubate at 35-37°C in 5-10% CO₂ for 18-24 hours till growth appears. Observe the zone of inhibition around the disc

Cultural response: On Tryptose Blood agar plates at 35-37°C for 18-24 hours

Organism(ATCC)	Diameter of zone of Inhibition
Streptococcus pyogenes (19615)	15 mm.

RESULTS:

A zone of inhibition \geq 14 mm. indicates susceptibility to Bacitracin and is presumptive of group A streptococci. For further confirmation serological grouping can be performed.

Precautions : For Laboratory Use.

USE: For identification and differentiation of Group A streptococci(especially S.pyogenes) from other beta haemolytic streptococci.

Precautions :For Laboratory Use.

Storage:Between 2 to 8°C. For prolonged use store below -20°C

Packing: 25 discs in vial

Disclaimer:

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications. The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.