

TECHNICAL SHEET

BDD024

KOVAC'S REAGENT STRIPS

SPECIFICATIONS:

APPEARENCE: Paper strips of size 60mm×5 mm

Directions:

Indole production by the organism is observed by inserting the Kovac's reagent strip between the plug and inner wall of the tube above the inoculated Peptone water (B035) and incubated at 35-37 °C for 18-24 hours.

Cultural response: In Peptone Water at 35-37°C for 18-24 hours.

Organism (ATCC)	Indole Production
Escherichia coli (25992)	Positive
Enterobacter aerogenes (13048)	Negative

RESULTS:

Pink coloration of lower part of the strip is positive reaction and means microorganisms possess tryptophan activity. A negative reaction shows no color change.

Precautions : For Laboratory Use.

USE: Detection of Indole production by microorganisms

Storage: Between 2 to 8°C.

Packing :25 strips per vial

Disclaimer:

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications.

The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.