

BDD029

ARGININE HYDROCHLORIDE DISCS

SPECIFICATIONS:

APPEARANCE: Paper discs of 6 mm diameter

Directions:

1. To determine Arginine hydroxylation, the Arginine disc (BDD029) is added in the Moeller Decarboxylase Broth Base (B161) which is used as a negative control for studying hydroxylation or as a base for the addition of amino acids.
2. The test organism is inoculated into the broth containing the Arginine disc (BDD029).
3. The inoculated tubes are overlaid with sterile mineral oil and incubated at 35-37°C for up to 4 days.
4. A purple colour indicates the Arginine hydroxylation.

Cultural response:

Cultural characteristics observed in Moeller Decarboxylase Broth Base (B161) with added Arginine Hydrochloride discs (BDD029) after an incubation at 35-37°C up to 4 days (Inoculated tubes are overlaid with sterile mineral oil).

Organism (ATCC)	Arginine decarboxylation
Citrobacter freundii (8090)	Variable reaction
Enterobacter aerogenes (13048)	Negative reaction, yellow colour
Escherichia coli (25922)	Variable reaction
Klebsiella pneumoniae (13883)	Negative reaction, yellow colour
Proteus mirabilis (25933)	Negative reaction, yellow colour
Proteus vulgaris (13315)	Negative reaction, yellow colour
Pseudomonas aeruginosa (9027)	positive reaction, purple colour
Salmonella Paratyphi A (9150)	Delayed positive reaction/ positive reaction, purple colour
Salmonella Typhi (6539)	Delayed positive reaction/ positive reaction, purple colour
Serratia marcescens (8100)	Negative reaction, yellow colour
Shigella dysenteriae (13313)	Delayed positive reaction/ positive reaction, purple colour
Shigella flexneri (12022)	Delayed positive reaction/ positive reaction, purple colour
Shigella sonnei (25931)	Variable reaction

Precautions : For Laboratory Use.

USE:Arginine hydrochloride discs are used for Arginine hydrolysis test.

Storage: 2-8°C

Packing: 25 discs per vial

BIOMARK Laboratories-INDIA

www.biomarklabs.com

TECHNICAL SHEET

Disclaimer:

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications.

The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.

Page 02 of 02