

TECHNICAL SHEET

PP015	Mueller Hinton Agar No. 2 Plate	
Formula		
Ingredients:	gms/lit.	
Casein acid hydrolysate	17.50	
Meat heart infusion#	2.00	
Starch, soluble	1.50	
Agar	17.00	
#-Equivalent to Beef heart infusion		
Final pH (at 25°C): 7.3± 0.2		
Directions:		
Label the ready to use plate (PP015).Either streak, inoculate or surface spread the test inoculum (50-100 CFU) aseptically on the plate.		
Principle:		
Meat, infusion fromand casein acid hydrolysate provide nitrogenous compounds, carbon, Sulphurand other essential nutrients.Starch is added to absorb any toxic substances present in the medium.Agar is the solidifying agent.The Kirby-Bauer procedure is based on agar diffusion of antimicrobial substances impregnated on paper discs.This method employs disc with a single concentration of antimicrobial agent and the zone diameters observed are correlatedwith minimum inhibitory concentration (MIC) values.Different factors influence the disc diffusion susceptibility tests as, inoculum concentration,agar depth, disc potency, medium pH and beta – lactamase production by test organisms. A standardized suspension of the organism is swabbed over the entire surfaceof the medium. Paper discs impregnated with specific amounts of antimicrobial agents are then placed on the surface of themedium, incubated and zones of inhibition around each disc are measured. The susceptibility is determined by comparingwith CLSI standards.		
(I) QC Tests		
pH:	7.3 ± 0.2	
Color:	Light Amber coloured medium	
Appearance:	Sterile Mueller HintonNo.2Agar in 85mm disposable plates.	
(II) Sterility test		
Passes release criteria		
(III) Q.C. Test Microbiological		
Cultural characteristics observed after incubation at 35-37°C for 18-24 hours.		
MICROORGANISM (ATCC)	GROWTH	
Escherichia coli 25922	Luxuriant	
Staphylococcus aureus25923	Luxuriant	
Pseudomonas aeruginosa27853	Luxuriant	
Enterococcus faecalis 29212	Luxuriant	
Neisseria gonorrhoeae49226	Luxuriant	

Refer disclaimer Overleaf

TECHNICAL SHEET

Precautions :	1. In Vitro diagnostic use only.
	2. Read the label before opening the container
Limitations :	1. Since the nutritional requirements of organisms vary, some strains may be encountered that fail to grow or grow poorly on this medium.
Use:	For testing susceptibility of common and rapidly growing bacteria using antimicrobial discs by the Kirby-Bauer Method
Storage:	Store between 15-25°C. Use before expiry date on the label.
Packing:	20/50 disposable plates.

Disclaimer:

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related BIOMARKLABORATORIES publications.

The information contained in this publication is based on our in-house studies and market performance and is to the best of our knowledge true and accurate. BIOMARK LABORATORIES reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.